

# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20150921-MH46340  
**Report Reference** MH46340-20120501  
**Issue Date** 2015-SEPTEMBER-21

**Issued to:** NESSCAP CO LTD  
17 Dongtangiheung-ro 681beon-gil  
Giheung-gu  
Yongin-si Gyeonggi-do 17102 KOREA

**This is to certify that  
representative samples of**

COMPONENT - CAPACITORS, ELECTROCHEMICAL  
Models ESHLR-0090C0-002R7, ESLSR-0120C0-002R3,  
ESHSR-0100C0-002R7, ESHSR-0360C0-002R7A,  
ESHSR-0360C0-002R7A1

Have been investigated by UL in accordance with the  
Standard(s) indicated on this Certificate.

**Standard(s) for Safety:** UL 810A, Electrochemical Capacitors.

**Additional Information:** See the UL Online Certifications Directory at  
[www.ul.com/database](http://www.ul.com/database) for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's  
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance  
capabilities and are intended for use as components of complete equipment submitted for investigation rather  
than for direct separate installation in the field. The final acceptance of the component is dependent upon its  
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please  
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

